

# 计量标准考核证书

Certificate for Examination of Measurement Standard

[ 2021 ] 揭 量 标 法 证 字 第 445200-007 号

根据《中华人民共和国计量法》，按照《计量标准考核规范》的要求，考核合格，特发此证。

This is to certify that the measurement standard conforms with the requirements of the "Rule for the Examination of Measurement Standard" according to the "Law on Metrology of the People's Republic of China".

建标单位 广东省揭阳市质量计量监督检测所

Possessor of the Measurement Standard

计量标准名称 医用热力灭菌设备温度计校准装置

代码 0411900

Name of Measurement Standard

Code

测量范围 (-80~150) °C

Measuring Range

不确定度或准确度等级

$U=0.15^{\circ}\text{C}$  ( $k=2$ )

或最大允许误差

Uncertainty/ Accuracy Class/

Maximum Permissible Error

保存地点 广东省揭阳市质量计量监督检测所电热室

Installed in

发证机关(印章)

The Issuing Authority



发证日期 2021 年 04 月 02 日

Date Issued

有效期至 2026 年 04 月 01 日

Date of Expiry

|                                                                                                |                     |                            |                                                                                    |                                                                                                 |                                                   |
|------------------------------------------------------------------------------------------------|---------------------|----------------------------|------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------|---------------------------------------------------|
| 计<br>量<br>标<br>准<br>器<br><br>Measurement<br>Standard                                           | 名称<br>Name          | 型号<br>Model / Type         | 测量范围<br>Measuring<br>Range                                                         | 不确定度或准确度<br>等级或最大允许误差<br>Uncertainty/ Accuracy Class/<br>Maximum Permissible Error              | 制造厂及<br>出厂编号<br>Manufacturer and<br>Series number |
|                                                                                                | 无线温度<br>记录器         | TL-2                       | (-80~150) °C                                                                       | $U=0.15^{\circ}\text{C} \quad (k=2)$                                                            | 汕头市迪威谱科技<br>有限公司<br>/47142/47049/4702<br>0/47008  |
| 主<br>要<br>配<br>套<br>设<br>备<br><br>Main Auxiliary<br>Equipment                                  |                     |                            |                                                                                    |                                                                                                 |                                                   |
| 可<br>开<br>展<br>检<br>定<br>及<br>校<br>准<br>项<br>目<br><br>Verification and<br>Calibration<br>Items | 名称<br>Name          | 测量范围<br>Measuring<br>Range | 不确定度或准确度<br>等级或最大允许误差<br>Uncertainty/ Accuracy Class/<br>Maximum Permissible Error | 依据的计量检定规程<br>或技术规范的代号及名称<br>Verification Regulation or<br>Technical Specification and It's code |                                                   |
|                                                                                                | 医用热力<br>灭菌设备<br>温度计 | 室温~150°C                   | $\pm 0.5^{\circ}\text{C}$ 及以下                                                      | JJF 1308-2011《医用热力灭菌设备<br>温度计校准规范》                                                              |                                                   |