

计量标准考核证书

Certificate for Examination of Measurement Standard

[2021] 粤 量标 揭法 证字 第 116 号

根据《中华人民共和国计量法》，按照《计量标准考核规范》的要求，考核合格，特发此证。

This is to certify that the measurement standard conforms with the requirements of the "Rule for the Examination of Measurement Standards" according to "the Law on Metrology of the People's Republic of China".

建标单位名称 广东省揭阳市质量计量监督检测所
Possessor of the Measurement Standard

计量标准名称 箱式电阻炉校准装置
Name of Measurement Standard

代码 04119013
Code

测量范围 (500~1000) °C
Measuring Range

不确定度或准确度等级 $U=1.0(k=2)$
或最大允许误差

Uncertainty or Accuracy Class or
Maximum Permissible Error

保存地点 广东省揭阳市榕城区揭阳大道东质监检测综合楼二楼电热长度室
Installed in

发证机关 (印章)
The Issuing Authority

发证日期 2021
Date of Issue

有效期至 2026
Date of Expiry



名 称 Name	型 号 Type	测量范围 Measuring Range	不确定度或准确度等级 或最大允许误差 Uncertainty or Accuracy Class or Maximum Permissible Error		制造厂及 出厂编号 Manufacturer and Series Number
计量 标准器 Measurement Standard	温湿度试验设备自 动检定系统 Vtest-1101X	(500~1000) °C	$U=1.0 (k=2)$		微测仪器(福州) 有限公司 /20071515
主要 配套设备 Main Auxiliary Equipment	/	/	/		///
开 展 的 检 定 或 校 准 项 目 Verification or Calibration Items	名 称 Name	测量范围 Measuring Range	不确定度或准确度等级 或最大允许误差 Uncertainty or Accuracy Class or Maximum Permissible Error		依据的计量检定规程 或计量技术规范的编号及名称 Verification Regulation or Measurement Technical Specification and Its Code
	箱式电阻炉	500°C~1000°C	炉温均匀度: $U=1.5\text{ }^{\circ}\text{C}$ ($k=2$); 炉温偏 差: $U=1.2\text{ }^{\circ}\text{C}$ ($k=2$); 炉温 稳 定 度: $U=0.8\text{ }^{\circ}\text{C}$ ($k=2$)		JJF 1376-2012 《箱式电阻炉校准 规范》